

## QRM-CTDI-CTWater<sup>®</sup>


## A WATER-equivalent dosimetry phantom, traced back on the standards for CTDI evaluation in Computed Tomography.

QRM-CTDI-CTWater<sup>®</sup> dosimetry phantom is manufactured from CTWater<sup>®</sup>, a proprietary resin which exhibits the same x-ray attenuation characteristics as liquid water in the range of 80 kV to 140 kV tube voltage.


The phantom is constructed in a modular fashion with a smaller inner cylinder which tightly fits into a larger outer cylinder. Both cylinders offer four through holes near the perimeter to accomodate standard CT dose probes. Supplied are nine massive rods, made from CTWater<sup>®</sup>, to plug the holes not in use, and adapter plugs to house diverse pencil shaped ion chambers. Further adapter plugs can be manufactured on customers specification.

## Specifications

Base material	CTWater®
diameter body	. 32 cm
diameter head	16 cm
depth	.15 cm
hole positions according to [2, 3]	
A seperate plug housing yo	ur prefered ion
chamber will be included!	



CTDI-Dosimetry-Phantom



Schematic view of the phantom

## References

- [1] Evaluation and comparison of water-equivalent CTDI dosimetryphantoms
  Erb, J.; Schmidt, M.; Schmidt, B.; Kalender, W.A.
  Engineering in Medicine and Biology Society, 2000. Proceedings of the 22nd Annual International Conference of the IEEE
  Volume 1, Issue , 2000 Page(s):102 105
- [2] Department of Health and Human Services, Food and Drug Administration. 21 CFR Part 1020: Diagnostic x-ray systems and their major components; amendments to performance standard; Final rule. Federal Register, 49, 171 (1984)
  [3] IEC 1223-2-6
  - Evaluation and routine testing in medical imaging departments. Part 2-6: Constancy tests X-ray equipment for computed